

**Zuckerberg San Francisco General Hospital
Point of Care Testing Temperature Record
For Room Temperature Use**

Month:	Year:	Location
---------------	--------------	-----------------

Day	Room Temp (*16-25 °C)	Initials	Comments
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			
11			
12			
13			
14			
15			
16			
17			
18			
19			
20			
21			
22			
23			
24			
25			
26			
27			
28			
29			
30			
31			

Note: Glucose supplies temperature range is *16 – 30°C.

Reviewed by: _____

*NOTE: When temperature is outside of defined tolerance limits, corrective actions must be taken and may include, but is not limited to, taking another reading after two hours, submitting a work order to Facilities, and moving contents to another acceptable storage location while problem is being addressed. **Always notify unit supervisor/POCT and document all corrective actions taken.**