

**Zuckerberg San Francisco General Hospital
Point of Care Testing Temperature Record
For Room Temperature and Humidity Use**

Month:		Year:		Location
Day	Room Temp (*16-25 °C)	Rel Humidity (*5 – 95%) ok? (v=yes)	Initials	Comments
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				
31				

Note: Glucose supplies temperature range is *16 – 30°C.

Reviewed by: _____

*NOTE: When temperature or humidity is outside of defined tolerance limits, corrective actions must be taken and may include, but is not limited to, taking another reading after two hours, submitting a work order to Facilities, and moving contents to another acceptable storage location while problem is being addressed. **Always notify unit supervisor/POCT and document all corrective actions taken.**