

Printed Employee Name: _____

Location: _____

AVOXIMETER 1000E Problem Solving Test

1. What is the proper storage temperature range for the AVOXimeter 1000E cuvettes?
 - a. Frozen
 - b. Refrigerated
 - c. Room Temperature

2. Optical Quality Control Filters should be run _____ and Liquid Quality Control should be run _____ prior to patient testing.
 - a. Daily and daily.
 - b. Twice a day and daily.
 - c. Daily and weekly.
 - d. Weekly and daily.

3. Cuvettes stored improperly (not in a sealed bag with desiccant pouch) may:
 - a. Be used for Optical and Liquid Quality Control testing.
 - b. Be used for patient samples only.
 - c. Cause erroneous QC and patient results.
 - d. Be placed in the refrigerator overnight to restore their integrity.

4. The differently-colored optical filters:
 - a. Can only be used with AVOXimeter 1000E instruments of the same serial number.
 - b. Can be shared with any AVOXimeter 1000E instruments.
 - c. Are used for hemolyzed blood specimens.
 - d. Are used with liquid quality controls run out.

5. True or False: Fill the cuvette until the vent patch begins to bulge.



YOU MUST SCORE 100% TO PASS.
If you did not pass, please review the Procedure
and the Operator's Manual and retake the exam.

SCORE

Employee Signature / Date	Trainer Signature / Date