

ANALYTE	Level 1 Reagent Results		
	Visual	CLINITEK Status®+	CLINITEK Advantus®
pH	6.0 - 7.0	6.0 - 7.5	6.0 - 7.5
Protein	Negative	Negative	Negative
Glucose	Negative	Negative	Negative
Ketones	Negative	Negative	Negative
Bilirubin	Negative	Negative	Negative
Blood (Hemoglobin)	Negative	Negative	Negative
Nitrite	Negative	Negative	Negative
Urobilinogen	0.2 - 1 EU/dL	0.2 - 1.0 EU/dL	0.2 - 1.0 EU/dL
Specific Gravity	1.005 - 1.020	≤ 1.005 - 1.025	≤ 1.005 - 1.020
Leukocytes Esterase	Negative	Negative	Negative
Creatinine	100 - 300 mg/dL	100 - 300 mg/dL	100 - 300 mg/dL
Albumin	10 mg/L	≤ 10 mg/L	≤ 10 mg/L
A:C	< 30 mg/g Normal <small>(For Microalbustix® only)</small>	< 30 mg/g Normal <small>(For CLINITEK® Microalbumin 2 & CLINITEK® Microalbumin 9)</small>	< 30 mg/g Normal <small>(For CLINITEK Microalbumin 9)</small>
P:C	Normal <small>For Multistix PRO® 10 LS & Multistix PRO® 11)</small>	Normal <small>For Multistix PRO 10 LS & Multistix PRO 11)</small>	Normal <small>For Multistix PRO 10 LS & Multistix PRO 11)</small>

QC PHYSICAL CHARACTERISTICS

Property	Value Range	Method
Appearance	Clear	Visual
Color	Yellow - Green	Visual
Specific Gravity	1.003 - 1.007	Refractometer (Room Temp.)
pH	6.0 - 7.0	pH meter

PREGNANCY TEST

Method	Result
Siemens CLINITEST® hCG	Positive

ANALYTE	Level 2 Reagent Results		
	Visual	CLINITEK Status®+	CLINITEK Advantus®
pH	7.0 – 8.0	7.0 – 8.5	7.0 – 8.5
Protein	Trace - ≥ 300 mg/dL	Trace - ≥ 300 mg/dL	Trace - ≥ 300 mg/dL
Glucose	100 - 2000 mg/dL	100 - ≥ 1000 mg/dL	100 - ≥ 1000 mg/dL
Ketones	5 - 160 mg/dL	Trace - ≥ 160 mg/dL	Trace - ≥ 80 mg/dL
Bilirubin	1 + - 3 +	Small - Large	Small - Large
Blood (Hemoglobin)	Trace - 3 +	Trace - Lysed - Large	Trace - Lysed - Large
Nitrite	Positive	Positive	Positive
Urobilinogen	2 - 8 EU/dL	2.0 - ≥ 8.0 EU/dL	2.0 - ≥ 8.0 EU/dL
Specific Gravity	1.015 - 1.030	1.020 - ≥ 1.030	1.015 - ≥ 1.030
Leukocytes Esterase	Trace - 3 +	Trace - Large	Trace - Large
Creatinine	100 - 300 mg/dL	100 - 300 mg/dL	100 - 300 mg/dL
Albumin	30 - 150 mg/L	30 - 150 mg/L	30 - 150 mg/L
A:C	30 - 300 mg/g Abnormal - > 300 mg/g High Abnormal <small>(For Microalbustix® only)</small>	30 - 300 mg/g Abnormal - > 300 mg/g High Abnormal <small>(For CLINITEK® Microalbumin 2 & CLINITEK® Microalbumin 9)</small>	30 - 300 mg/g Abnormal - > 300 mg/g High Abnormal <small>(For CLINITEK® Microalbumin 9)</small>
P:C	Abnormal <small>For Multistix PRO® 10 LS & Multistix PRO® 11)</small>	150 - > 500 mg/g Abnormal <small>For Multistix PRO 10 LS & Multistix PRO 11)</small>	150 - > 500 mg/g Abnormal <small>For Multistix PRO 10 LS & Multistix PRO 11)</small>

QC PHYSICAL CHARACTERISTICS

Property	Value Range	Method
Appearance	Clear - Slightly Cloudy	Visual
Color	Amber - Red	Visual
Specific Gravity	1.025 - 1.029	Refractometer (Room Temp.)
pH	7.0 - 8.0	pH meter

PREGNANCY TEST

Method	Result
Siemens CLINITEST® hCG	Negative